

DOCUMENT RESUME

ED 118 891

CE 006 450

AUTHOR Glorig, Aram, Ed.; And Others
TITLE Guidelines for a Training Program for Audiometric Technicians. Report of Working Group 66.
INSTITUTION National Academy of Sciences - National Research Council, Washington, D.C. Committee on Hearing, Bioacoustics, and Biomechanics.
SPONS AGENCY Office of Naval Research, Arlington, Va.
PUB DATE 19 Sep 73
NOTE 17p.

EDRS PRICE MF-\$0.83 HC-\$1.67 Plus Postage
DESCRIPTORS *Audiometrists; Auditory Tests; Auditory Training; Course Content; *Course Descriptions; Ears; Educational Programs; Guidelines; *Hearing Conservation; *Lesson Plans; Technical Education; Training Techniques

ABSTRACT

The document outlines a course designed to train audiometric technicians who will conduct pure-tone conduction tests as part of a program on hearing conservation in noise. A minimum of two days is required for the completion of the course. The outline of the training program presents nine topics with an indication of the minimum time required for each topic. The topics include: (1) hearing conservation in noise (importance of hearing, objectives of training program, and the audiometric technician's responsibilities and limitations); (2) basic discussion of sound and its measurement (parameters of sound and definitions and description and use of sound measuring equipment); (3) basic anatomy and physiology of the ear (function of the three parts of the ear and disorders of the ear); (4) the audiometer (type, how it works, and calibration and care); (5) audiometric technique (instruction to subject and procedure, record keeping, audiometric scheduling, pitfalls to avoid, and supervised practicum); (6) review of practicum and further audiograms; (7) medico-legal aspects and films (compensation, discussion, and films); (8) hearing protection and fitting procedures (advantages of muffs and plugs and practicum in fitting procedures); and (9) examination and review. (Author/EC)

* Documents acquired by ERIC include many informal unpublished *
* materials not available from other sources. ERIC makes every effort *
* to obtain the best copy available. Nevertheless, items of marginal *
* reproducibility are often encountered and this affects the quality *
* of the microfiche and hardcopy reproductions ERIC makes available *
* via the ERIC Document Reproduction Service (EDRS). EDRS is not *
* responsible for the quality of the original document. Reproductions *
* supplied by EDRS are the best that can be made from the original. *

ED118891

GUIDELINES FOR A TRAINING PROGRAM FOR AUDIOMETRIC TECHNICIANS

01006450

U S DEPARTMENT OF HEALTH,
EDUCATION & WELFARE
NATIONAL INSTITUTE OF
EDUCATION

THIS DOCUMENT HAS BEEN REPRO-
DUCED EXACTLY AS RECEIVED FROM
THE PERSON OR ORGANIZATION ORIGIN-
ATING IT. POINTS OF VIEW OR OPINIONS
STATED DO NOT NECESSARILY REPRESENT
OFFICIAL NATIONAL INSTITUTE OF
EDUCATION POSITION OR POLICY.

NAS-NRC Committee on Hearing, Bioacoustics, and Biomechanics

WORKING GROUP 66

Report of Working Group 66,
Guidelines for a Training Program
for Audiometric Technicians.
NAS-NRC Committee on Hearing, Bio-
acoustics, and Biomechanics, 1973.

NATIONAL ACADEMY OF SCIENCES - NATIONAL RESEARCH COUNCIL
COMMITTEE ON HEARING, BIOACOUSTICS, AND BIOMECHANICS

GUIDELINES FOR A TRAINING PROGRAM FOR AUDIOMETRIC TECHNICIANS

Report of Working Group 66
Aram Glorig, Chairman

William L. Baughn
G.R. Bearce
Marion Downs
Donald C. Gasaway
Walter H. Konn

Roger Maas
Rufus W. Miller
Joseph Sataloff
B. Thomas Sheib
Douglas E. Wheeler

Office of Naval Research
Contract No. N00014-67-A-0244-0021

Reproduction in whole or
in part is permitted for
any purpose of the United
States Government.

September 1973

NOTICE

The project which is the subject of this report was approved by the Governing Board of the National Research Council, acting in behalf of the National Academy of Sciences. Such approval reflects the Board's judgment that the project is of national importance and appropriate with respect to both the purposes and resources of the National Research Council.

The members of the committee selected to undertake this project and prepare this report were chosen for recognized scholarly competence and with due consideration for the balance of disciplines appropriate to the project. Responsibility for the detailed aspects of this report rests with that committee.

Each report issuing from a study committee of the National Research Council is reviewed by an independent group of qualified individuals according to procedures established and monitored by the Report Review Committee of the National Academy of Sciences. Distribution of the report is approved, by the President of the Academy, upon satisfactory completion of the review process.

9/19/73

Foreword

This course is designed to train audiometric technicians who will conduct pure-tone air conduction tests as part of a program of Hearing Conservation in Noise. A minimum of two days is required for the completion of the course, and a minimum requisite time for each of the topics is indicated in the course outline.

Two basic readings should accompany the course: "An Approved Technique for Pure Tone, Air Conduction Audiometry" (Glorig, A., Chairman, CHABA Working Group 66), Appendix, and "A Guide for Industrial Audiometric Technicians (Am. Industrial Hygiene Association, 14125 Prevost, Detroit, Michigan 48227).

Other readings that are useful as background material for the course are:

1. "Guide for Conservation of Hearing in Noise," Subcommittee on Noise in Industry, Aram Glorig, M.D., 1966 Inwood Road, Dallas, Texas 75235.
2. Industrial Noise Manual, (\$15.30), Am. Industrial Hygiene Association, 14125 Prevost, Detroit, Michigan 48227.
3. Glorig, A., Noise and Your Ear, Greene and Stratton, Inc., 1958.
4. Sataloff, J., Hearing Loss, Lippincott, New York, 1966.
5. "Guide for Industrial Audiometric Technicians," Employers Mutuals of Wausau, Wausau, Wisconsin, 1967.
6. "Industrial Noise and Hearing Protection," Employers Mutuals of Wausau, Wausau, Wisconsin.

In addition, pamphlets and brochures on audiometry are often provided by Audiometer Manufacturers and can be obtained on request.

CONTENTS

I. Hearing Conservation in Noise -- Introductory	1
II. Basic Discussion of Sound and Its Measurement	2
III. Basic Anatomy and Physiology of the Ear	3
IV. The Audiometer	3
V. Audiometric Technique	5
VI. Review of Practicum and Further Audiograms	8
VII. Medico-Legal Aspects and Films	8
VIII. Hearing Protection and Fitting Procedures	9
IX. Examination and Review	9

SPECIFICATIONS ON A TRAINING PROGRAM FOR AUDIOMETRIC
TECHNICIANS IN HEARING CONSERVATION IN NOISE

F I R S T D A Y

- TOPIC 1. Hearing Conservation in Noise -- Introductory
Statement (No less than 45 minutes)
- A. Importance of Hearing
1. Social Aspects
- a. The Meaning of Hearing Loss -- suggested
demonstrations, any of the following:
- (1) Zenith record: "Getting Through"
Zenith Radio Corporation, 6501 West Grand
Avenue, Chicago, Illinois 60635
- (2) Harford record: "How They Hear" --
Gordon Stowe and Association, P.O. Box 233,
Northbrook, Illinois
- (3) Tape "Demonstration of Hearing Impair-
ment" -- Price Filmmakers, Inc., 3491 Cahuenga
Boulevard, West, Suite 2, Hollywood, Cali-
fornia 90028.
2. Economic and Legal Aspects
- B. Objectives of Training Program
1. Valid Audiograms -- Medical, legal, and social
reasons why a valid audiogram is requisite
2. Effective Ear Protection
3. Medical Follow-up

C. The Audiometric Technician's Responsibilities and Limitations: Should be able to produce a valid audiogram, to recognize when it is or is not valid, and to retake it when necessary. The physician only is to interpret the audiogram and to make a diagnosis from it.

TOPIC II. Basic Discussion of Sound and Its Measurement

(No less than 45 minutes)

A. Parameters of Sound and Definitions

1. Frequency (Contrast with Pitch)

2. Intensity (Contrast with Loudness)

a. Meaning of the Decibel and Attenuation

b. Three types of Intensity Measures

(1) SPL (Sound Pressure Level)

(2) HL (Hearing Level) -- ASA

(3) HL (Hearing Level) -- ANSI

B. Description and Use of Sound Measuring Equipment

(Background information only, not intended to qualify for making noise measurements)

1. Sound Level Meter

a. Measurements

(1) "dBA" -- Its meaning and why it was chosen

(2) B scale

(3) C scale

3.

2. Why Hazard is Associated with "dBA" and the Percentage Risk
3. Noise as a Masker of Signals
 - a. Sound-treated booth is the only acceptable environment for threshold testing because of the masking effect of background noise on pure tones
 - b. Sound-room specifications by ANSI can be checked by reference to manufacturer's tests
 - c. The audiometric technician must assume responsibility for making sure the test booth is used properly and inside ambient level is checked periodically.

TOPIC III. Basic Anatomy and Physiology of the Ear (No less than 45 minutes)

- A. Function of the Three Parts of the Ear
 1. Outer
 2. Middle -- Including Eustachian Tube Function
 3. Inner
- B. Disorders of the Ear
 1. Conductive Disorders and Audiograms
 2. Sensorineural Disorders and Audiograms
 3. Types of Audiograms and Purposes
 - a. Screening Audiograms
 - b. Diagnostic

TOPIC IV. The Audiometer (No less than 90 minutes)

- A. Types
 1. Manual
 2. Automatic
- B. How it Works -- Functional Parts of Audiometer Common to All

C. Calibration and Care

1. Daily -- Self-Measurement (Check audiometer on own ear)
 2. Monthly -- Biological Calibration is obtained by testing 5 or 6 young female subjects (under 25) who will be available for recheck. Audiometer should be initially checked on this group when received, and the results recorded. Once a month, 2 or 3 of this group should be retested. If the average of their thresholds is 10 dB or more off at one frequency, or 5 dB off at three frequencies, the audiometer should be sent for physical calibration. (Don't make corrections for the difference.)
 3. Yearly -- Physical Calibration should be obtained once a year, by a properly qualified service organization. Audiometer should be returned for calibration complete with earphone assembly, etc.
4. Care
- a. Earphone Care
 - (1) Gentleness in Handling
 - (2) Clean Cushions with soap and water, lightly
 - b. Audiometer should be turned on in morning and left on all day. Do not switch on and off--it is harmful to the instrument

c. Trouble shooting check points

- (1) Broken Cords
- (2) Loose Phone Cushions
- (3) Dirt in Grid of Earphone
- (4) Tension of Headband
- (5) Noise in Attenuator (listen to sound during dial changes)
- (6) Self-recording audiometers -- The peak-to-peak excursions must be less than 15 dB to be valid

TOPIC V. Audiometric Technique (No less than 45 minutes, plus 2 hour practicum)

A. Instruction to Subject and Procedure -- Demonstrating Technique

B. Record Keeping

1. Recording Results by means of serial audiogram forms (Record thresholds roughly and then copy to serial; or cover up previous test results when recording)

2. Self-recording Tests -- Take last half of tracing, draw a line halfway between peaks and record to closest 5 dB

C. Audiometric Scheduling

1. "Base-line" Audiogram -- A baseline audiogram is one obtained at initial employment time, or at a time when the ears have been away from noise 14 to 48 hours.

These should be scheduled as follows:

- a. Pre-Employment
 - b. When a Change of 15 dB or more occurs at any frequency
 - c. At Intervals designated by physician (e.g., every 6 years)
2. Routine Audiometric Monitoring -- Audiograms should be obtained at regular intervals (e.g., every two years), as specified by the physician
- D. Pitfalls To Avoid
1. Subjects Observing the manipulation of the dials (seat at an angle)
 2. Visual Clues such as facial expressions or eye movements by the tester
 3. Poor placement and adjustment of headphones
 - a. Remove Obstructions such as glasses, hair, earrings; headphones should be placed and removed only by tester; subject should be cautioned not to touch phones
 - b. Earphone grid should be opposite ear aperture
 4. Tinkering with headband pressure
 5. Vague Instructions -- Give clear directions to respond to faintest sound
 6. Rhythmic Presentation of tone -- Vary both the inter-signal interval and the duration of tone presentation
 - a. No more than 2-1/2 seconds
 - b. No less than 1 second

7.

7. Incomplete forms (Be sure to record all information such as age, sex, date, etc.)

8. Don't Turn Audiometer on and off -- leave on all day

9. Problems with tinnitus -- Break rhythm and pulse tones to avoid confusion with head noise

10. Noise Problems in Testing

a. If too many audiograms show drops at 500 Hz, or if jack panels on booth are loose, or if seals appear to be broken in booth, inform manufacturer. Seals should be inspected every 6 months

b. Instrument should be placed outside testing booth

11. Cord Interference -- Cords should be free from contact with subject

12. Over-long Testing -- 5-8 minutes is sufficient for test. If it takes longer than this, call back for later recheck. If suspicion of malingering, no hint should be given and patient should be referred

13. Interrupter Switch -- leave on off position on dial

E. Supervised Practicum -- (no less than 2 hours) -- Students are paired and rotate preferably no more than two to an audiometer. Instructors instruct up to 8 students who circulate among pairs. Both automatic and manual screening audiometers should be represented; these may be obtained by requesting students to bring their own

audiometers and/or inviting manufacturers to represent their instruments at the course. Approximately 6 complete audiograms can be done during this time.

TOPIC VI. Review of Practicum and Further Audiograms (No less than 2 hours)

A. At least another 6 audiograms should be made during this time

TOPIC VII. Medico-Legal Aspects and Films (No less than 90 minutes)

A. Compensation

1. Importance of audiogram base lines from which comparison can be made
2. Kinds of laws in various states
3. Formula for calculating percentage of loss

B. Discussion and Films (Films can be shown at this point, or they can be shown at other appropriate times during the courses as directed by the instructor, at his discretion)

1. Approved Films from Price Filmmakers, Inc.*
 - a. Hear: It Takes Two -- The importance of hearing to man
 - b. Audiometric Techniques -- Demonstration of methods of obtaining audiograms

* Price Filmmakers, Inc., 3491 Cahuenga Blvd., West, Suite 2, Hollywood, California 90028.

c. Ear Protection and Noise -- Personal ear protection for workers in high noise levels

d. Noise and Its Effects on Man -- Noise in Everyday Life and How it is Combated

TOPIC VIII. Hearing Protection and Fitting Procedures

(No less than 90 minutes)

A. Advantages of Muffs and Plugs

1. Cautions

a. Don't accept any that don't have an independent laboratory measurement report

b. "Amplitude Sensitive" devices are inappropriate for industrial use

B. Practicum in Fitting Procedures with Experience

Fitting Different Types of Plugs

TOPIC IX. Examination and Review (No less than 90 minutes)

A certificate will not be issued until each participant has submitted the results of ten tests on ten individuals, made at his testing location. This will ensure that an active testing program is going on, and that the trainee will be using his skills soon enough not to forget them.

A yearly refresher course of one day is recommended for each trainee.

DOCUMENT CONTROL DATA - R&D

(Security classification of title, body of abstract and indexing annotation must be entered when the overall report is classified)

1. ORIGINATING ACTIVITY (Corporate author) National Academy of Sciences - National Research Council Committee on Hearing, Bioacoustics, and Biomechanics	2a. REPORT SECURITY CLASSIFICATION None
	2b. GROUP None

3. REPORT TITLE

Guidelines for a Training Program for Audiometric Technicians

4. DESCRIPTIVE NOTES (Type of report and inclusive dates)

5. AUTHOR(S) (Last name, first name, initial)

Dr. Aram Glorig, editor

6. REPORT DATE September, 1973	7a. TOTAL NO. OF PAGES 13	7b. NO. OF REFS 6
---------------------------------------	----------------------------------	--------------------------

8a. CONTRACT OR GRANT NO. N00014-67-A-0244-0021 b. PROJECT NO. c. d.	9a. ORIGINATOR'S REPORT NUMBER(S)
	9b. OTHER REPORT NO(S) (Any other numbers that may be assigned this report)

10. AVAILABILITY/LIMITATION NOTICES

Qualified Requesters may obtain copies of this report from DDC.

11. SUPPLEMENTARY NOTES	12. SPONSORING MILITARY ACTIVITY Office of Naval Research Code 454 Arlington, Virginia 22217
-------------------------	---

13. ABSTRACT

This course is designed to train audiometric technicians who will conduct pure-tone air conduction tests as part of a program of Hearing Conservation in Noise. A minimum of two days is required for the completion of the course., and a minimum requisite time for each of the topics is indicated in the course outline.

14. KEY WORDS	LINK A		LINK B		LIN
	ROLE	WT	ROLE	WT	ROLE
audiometry hearing conservation hearing threshold audiometers ear protection					



14. KEY WORDS	LINK A		LINK B		LINK C	
	ROLE	WT	ROLE	WT	ROLE	WT
audiometry hearing conservation hearing threshold audiometers ear protection						

INSTRUCTIONS

1. ORIGINATING ACTIVITY: Enter the name and address of the contractor, subcontractor, grantee, Department of Defense activity or other organization (*corporate author*) issuing the report.

2a. REPORT SECURITY CLASSIFICATION: Enter the overall security classification of the report. Indicate whether "Restricted Data" is included. Marking is to be in accordance with appropriate security regulations.

2b. GROUP: Automatic downgrading is specified in DoD Directive 5200.10 and Armed Forces Industrial Manual. Enter the group number. Also, when applicable, show that optional markings have been used for Group 3 and Group 4 as authorized.

3. REPORT TITLE: Enter the complete report title in all capital letters. Titles in all cases should be unclassified. If a meaningful title cannot be selected without classification, show title classification in all capitals in parenthesis immediately following the title.

4. DESCRIPTIVE NOTES: If appropriate, enter the type of report, e.g., interim, progress, summary, annual, or final. Give the inclusive dates when a specific reporting period is covered.

5. AUTHOR(S): Enter the name(s) of author(s) as shown on or in the report. Enter last name, first name, middle initial. If military, show rank and branch of service. The name of the principal author is an absolute minimum requirement.

6. REPORT DATE: Enter the date of the report as day, month, year; or month, year. If more than one date appears on the report, use date of publication.

7a. TOTAL NUMBER OF PAGES: The total page count should follow normal pagination procedures, i.e., enter the number of pages containing information.

7b. NUMBER OF REFERENCES: Enter the total number of references cited in the report.

8a. CONTRACT OR GRANT NUMBER: If appropriate, enter the applicable number of the contract or grant under which the report was written.

8b, 8c, & 8d. PROJECT NUMBER: Enter the appropriate military department identification, such as project number, subproject number, system numbers, task number, etc.

9a. ORIGINATOR'S REPORT NUMBER(S): Enter the official report number by which the document will be identified and controlled by the originating activity. This number must be unique to this report.

9b. OTHER REPORT NUMBER(S): If the report has been assigned any other report numbers (*either by the originator or by the sponsor*), also enter this number(s).

10. AVAILABILITY/LIMITATION NOTICES: Enter any limitations on further dissemination of the report, other than those

imposed by security classification, using standard statements such as:

- (1) "Qualified requesters may obtain copies of this report from DDC."
- (2) "Foreign announcement and dissemination of this report by DDC is not authorized."
- (3) "U. S. Government agencies may obtain copies of this report directly from DDC. Other qualified DDC users shall request through _____."
- (4) "U. S. military agencies may obtain copies of this report directly from DDC. Other qualified users shall request through _____."
- (5) "All distribution of this report is controlled. Qualified DDC users shall request through _____."

If the report has been furnished to the Office of Technical Services, Department of Commerce, for sale to the public, indicate this fact and enter the price, if known.

11. SUPPLEMENTARY NOTES: Use for additional explanatory notes.

12. SPONSORING MILITARY ACTIVITY: Enter the name of the departmental project office or laboratory sponsoring (*paying for*) the research and development. Include address.

13. ABSTRACT: Enter an abstract giving a brief and factual summary of the document indicative of the report, even though it may also appear elsewhere in the body of the technical report. If additional space is required, a continuation sheet shall be attached.

It is highly desirable that the abstract of classified reports be unclassified. Each paragraph of the abstract shall end with an indication of the military security classification of the information in the paragraph, represented as (TS), (S), (C), or (U).

There is no limitation on the length of the abstract. However, the suggested length is from 150 to 225 words.

14. KEY WORDS: Key words are technically meaningful terms or short phrases that characterize a report and may be used as index entries for cataloging the report. Key words must be selected so that no security classification is required. Identifiers, such as equipment model designation, trade name, military project code name, geographic location, may be used as key words but will be followed by an indication of technical context. The assignment of links, roles, and weights is optional.