

**Project: IEEE P802.15 Working Group for Wireless Personal Area Networks (WPANs)**

**Submission Title:** [YNU's PHY and MAC design for WBAN IEEE P802.15.6]

**Date Submitted:** [3 May, 2009]

**Source:** [Ryuji Kohno, Koji Enda, Shun Nagamine, Hideki Mochizuki, Haruka Suzuki] Company  
[Yokohama National University]

Address [79-5 Tokiwadai, Hodogaya-ku, Yokohama, Japan 240-8501]

Voice:[+81-45-331-4115], FAX: [+81-45-331-1157], E-Mail:[kohno@ynu.ac.jp]

**Abstract:** [We propose our specially designed PHY and MAC for wireless body area network (BAN) to satisfy various requirements for both medical and non-medical applications, which is uniqueness of .IEEE802.15.6. In PHY, pulsed chirp UWB is proposed because of its total performance for WBAN. In particular, hybrid type of ARQ and FEC is proposed to satisfy both requirements of medical and non-medical applications in a sense of highly reliable and lower power consumption. In MAC, we propose a protocol for BAN considering effect to a human body, in particular, taking care of SAR (specific absorption ratio) of parts of BAN devices in and on a body. Although this proposal is not a full set of proposal corresponding to technical requirements for IEEE802.15.6, we hope this can contribute to improve system performance by harmonizing with others. This is a revised version of our proposal with a new positioning scheme in addition in March, 2009.]

**Purpose:** [Response to "TG6 Call for Proposals" (IEEE P802.15-08-0811-02-0006).]

**Notice:** This document has been prepared to assist the IEEE P802.15. It is offered as a basis for discussion and is not binding on the contributing individual(s) or organization(s). The material in this document is subject to change in form and content after further study. The contributor(s) reserve(s) the right to add, amend or withdraw material contained herein.

**Release:** The contributor acknowledges and accepts that this contribution becomes the property of IEEE and may be made publicly available by P802.15.

# **YNU's PHY and MAC Design for WBAN IEEE P802.15.6**

**- Revision of Our Proposal in March, 2009 -**

Ryuji Kohno, Koji Enda, Shun Nagamine,  
Hideki Mochizuki, Haruka Suzuki  
Yokohama National University

# 1. Aim and Motivation

- The aim of this presentation is to introduce some technologies to enhance performance of a merged proposal.
- We propose our designed schemes for PHY and MAC of BAN so as to satisfy the different requirements for both medical and non medical use of BAN such as reliability and safety for medical use and efficiency for non-medical use.
- Positioning or localization scheme for BAN nodes is also presented in addition different from the last presentation in March.

## 2. PHY

2.1 Pulsed Chirp UWB using hopping

2.2 Error controlling scheme using hybrid ARQ and FEC for medical and non medical uses, respectively

## 3. MAC

3.1 Protocol considering SAR or thermal influence to a body by switching cluster


3.2 Positioning or localization of BAN nodes

## 2. Physical Layer design: Pulsed Chirp UWB with Hopping

## 2. Body Area Network: BAN

### •Body Area Network

Networks composed of wireless communication inside and/or outside a human body




Implant WBAN

Wearable WBAN

1. Monitoring vital signs such as ECG, EEG etc and remotely controlling medical equipments such as pace-maker with ICD, capsule endoscope
2. Assisting disability such as eyeglasses with camera, etc.
3. Fitness and Entertainment with body area games and cell phones etc.

●:Wearable ●:Implant

Example of implant device




# 2.1 Direct Sequence UWB (DS-UWB) and Chirp on UWB (Co-UWB)

## DS-UWB system

### Characteristic

- advantage : transmitting signal for DS-UWB using spreading sequence
- down back : **no tolerance to near-far problem due to no spreading gain**


↪ We can distinguish many users



## Chirp on UWB (Co-UWB) system

### Characteristic

- advantage : Easily sweeping frequency among ultra wideband  
 ↪ Using different frequency band for multiple nodes, **tolerance to near-far problem**
- down back : **it is difficult to distinguish different BANs in case using the same frequency band**



## 2.1 Pulsed Chirp UWB system

Aim: Improve immunity against narrow band and multi-user interference with combined chirp-on-UWB and DS-UWB


Chirp-on-UWB system

+

DS-UWB system

Using frequency sweeping, it is good performance against narrow band interference in case SIR is low.

Using direct sequence, It is good performance against multi user interference in case SIR is high.



Proposed system:

**Pulsed Chirp UWB system**; like IEEE802.15.4a with optional chirp pulse  
This seems to be a merged proposal between NICT's Part 1&2\* in a sense of  
“**Multi-Band Pulsed Chirp UWB with Frequency and Time Hopping**”

Each pulse uses different sub-band for tolerance against narrow band interference

Transmitting signal uses spreading sequence to distinguish multiple nodes in BANs


\*: NICT's Part 1: Doc#IEEE802.25.09.0166-00-0006, NICT's Part 2: Doc#IEEE802.25.09.0163-00-0006,

# 2.1 Pulsed Chirp UWB system with single hopping for medical use

The system uses multiple sub-bands with OCC frequency hopping sequence to avoid both narrow band and multi-user interferences and DS with Gold sequence to indentify users.

This system seems to be a merged proposal between NICT's Part 1&2 in a sense of "Multi-Band Pulsed Chirp UWB with Frequency and Time Hopping"

- ① Divide the bandwidth by applied length of sequence
- ② Define using frequency sub-band which adapts to each pulse by hopping sequence
- ③ Make direct sequence by using direct sequence with up and down chirp corresponding 0 and 1.
- ④ Data modulation is bi-phase





# 2.1 Pulsed Chirp UWB system with parallel hopping for non-medical use


**Medical use** → Utilize hopping sequence for improvement of interference immunity



**Non-medical use** → Utilize hopping sequence for multi-level modulation for satisfying demand of higher data rate


### For medical use

$n_i(j)$  : j-th element of Hopping sequence assigned user i  
:user1 (blue) :user2 (green) :user3 (red)



### Tasks of Hopping sequence


- **For medical use:** Use for identification of multiple BAN nodes
- **For non Medical use:** Use for higher data-rate transmission with multi-level modulation




### For non-medical use

A non-medical user employs 3 parallel sub-bands to transmit data with 3 times higher data rate

# 2.1 System configuration for **medical use**



# 2.1 System configuration for non-medical use



# 2.1 Wearable Wireless BAN channel model

Using channel model of IEEE802.15.4a final report


### Characteristic

- frequency is 3.4~4.8GHz



No energy penetrating through a body

Path loss model are defined according to distance around a body




Path loss formula

$$P_{dB} = \gamma (d - d_0) + P_{0,dB} \quad 12$$

$\gamma$ : units of dB/meter      $d_0$ :reference distance

$P_{0,dB}$ :Power at the reference distance



Diffracting wave around the body


Ground reflection

Depending on transmitter and receiver position, the power of reflection wave is higher

Ex) Impulse response which is receiver's position left arm

## 2.1 Simulation Model

- Simulation model
  - Multiple pico-nets (Multiple BANs) coexisting
  - Assuming multiple BANs are perfectly synchronized and channel propagation is known in ideal case
  - Undesired factor : multi-user interfering signals and AWGN
- Parameter of each system
  - Total frequency sub-bands used in all BANs is the same in each system
  - Bit rate and power consumption of 1bit is the same
  - Number of multiple nodes is the same



# 2.1 Performance Evaluation


## Simulation Parameters

Both medical and non-medical nodes and BANs are coexisting.


<b>Pulse shape</b>	<b>Root raised cosine roll-off pulse( roll-off rate 0.6)</b>
<b>Bit rate</b>	<b>Medical:1Mbps Non-medical:3Mbps</b>
<b>Frequency Band</b>	<b>3.2 - 4.8GH</b>
<b>Sampling interval</b>	<b>0.08[ns]</b>
<b>Using channel model</b>	<b>IEEE 802.15.6 CM3</b>
<b>No of coexisting BAN nodes for medical and non-medical use</b>	<b>8</b>

Type of UWB	Pulse Duration
DS-UWB using Gold sequence with length 7	0.75ns
Proposed Pulsed Chirp UWB with FH and DS using RS code with length 8 as FH and Gold sequence with length 7 as DS	2.5ns~ 20ns

# 2.1 BER Performance Evaluation



Case of the number of coexisting nodes or BANS is 8 (Bit rate = 1Mbps)



Case of the number of coexisting nodes or BANS is 8 (Bit rate = 3Mbps)

## 2.1 Link budget

parameter	Value(medical)	Value(non-medical)
Data rate(R)	1Mbps	3Mbps
Tx antenna gain(Gt)	0dBi	0dBi
Rx antenna gain(Gr)	0dBi	0dBi
Required(SIR) for BER <sub>10</sub> <sup>-3</sup>	-16	-9
Received SNR(dB)	15	15
Path loss at 1m(dB)	60.98	60.98




## 2.1 Conclusions for Pulsed Chirp UWB

- We proposed pulsed chirp UWB with **single and parallel frequency hopping for medical and non-medical use, respectively.**
- The proposed pulsed chirp UWB with hopping sequence performs **stable high interference immunity in various environments of coexisting medical and non-medical nodes or BANs.**
- The proposed UWB may be a merged proposal between **NICT's Part 1&2 in a sense of "Multi-Band Pulsed Chirp UWB with Frequency and Time Hopping"**

## 2.2 Error Controlling Scheme in Physical Layer Design:

Combination between Hybrid ARQ for  
Medical Use and FEC for Non-Medical Use

## 2.2 Demand of medical and non medical system



### Medical

- Robust against interference
- High reliability and security

### Non-medical

- Efficiency with High rate
- Continuous data streaming




**Aim : Design an error-controlling scheme for BAN appropriate for both medical and non-medical uses**

### Coding for higher reliability



→ Degradation of throughput according to more redundancy for high error-correcting capability

### Multi-level modulation for high speed



Shorter Euclidian distance among signal points ● ●

→ Increase of error rate according to higher efficiency

**Solution: Choose decoding scheme between hybrid ARQ and simple FEC for medical and non-medical uses, respectively in reception while the same modulation and coding are used for both medical and non-medical uses in transmission**

# 2.2 Several types of Error controlling method

## FEC

(Forward Error Correction)

Error detection or error correction by adding redundant bit without retransmission

## ARQ

(Automatic Repeat reQuest)

If packet error is detected, it is retransmitted.

FEC+ARQ is

# Hybrid ARQ


Error detection and correction is performed in decoding at reception and if the error is detected and cannot be corrected, retransmission is requested until correctly received.

■ Type 1 of hybrid ARQ: Retransmit **same data**

■ Type 2 of hybrid ARQ: Retransmit **other data**


**Incremental Redundancy method : Retransmit redundant bit**

high **throughput** communication



Modulation: BPSK  
Channel: AWGN

## 2.2 Code Configuration (Interleaving)



## 2.2 Proposed combined hybrid ARQ and FEC for error-control satisfying different requirements for medical and non-medical applications

Choose decoding scheme either hybrid ARQ or simple FEC according to medical or non-medical use while transmitting signals are the same in transmission device

- Requirement for Medical use: Accept a certain level of delay for improvement of quality
- Requirement for Non-medical use : Decrease delay in moderate quality


### Proposed method

**For Medical use: Hybrid ARQ**  
**For Non-medical use: FEC only**

Use **Super orthogonal convolutional code** or **Concatenated code** according as a purpose


## 2.2 Flowchart of proposed error-controlling scheme

(1) Make a block with k bits by dividing binary data stream




(1) Add CRC code to detect errors  
Convolutional Coding

(2) B blocks of data are added with CRC and encoded with a coderate  $R_0$  of convolutional code.



(2) Initial packet L1  
(both Medical and Non-Medical uses)


(3) Make the first transmitting block L1 and retransmitting redundant block L2, L3 for medial use.



(3) Retransmission packet L2,L3  
(Medical use only)

## 2.2 Reason for Retransmission


2-PPM (2-level Pulse position modulation)



**Avoid the interference by expanding time duration  $T_d$**



## 2.2 Time Duration $T_d$





**Short duration  $T_d$  : Increase errors due to ISI**  
**Long duration  $T_d$  : Increase errors due to Noise**

## 2.2 Performance Evaluation for Proposed Error-Controlling Scheme

simulation Parameters

Channel	Multipath model considering on body transmission
Modulation	2PPM, Squared detection (soft decision)
Code	Coding rate: 1/3 RCPC Code : CRC16
Block number length	300 bit , 10 block
Decode	Hard decision of Viterbi decoding
Limit of number of retransmission	4
UWB pulse	Gaussian mono-cycle pulse
Pulse width	2ns
Symbol duration	4ns

## 2.2 Evaluation of Bit Error Rate and Throughput



- Medical use : BER=10<sup>-6</sup>
- Non medical use : BER=10<sup>-3</sup>

- Medical use : Max 1Mbps
- Non medical use : Max 10Mbps

## 2.2 Another Hybrid ARQ method (1)

### Chase Combining method : Retransmit same bit

$\mathbf{I} = I_1 \dots I_K$  : data (k bits)      effective for fading channels and jamming

$\mathbf{X} = X_1 \dots X_N$  : codeword (coderate  $K/N$ )

$\mathbf{Y}^1 = Y_1 \dots Y_N, w_1$

$\mathbf{Y}^L = Y_1 \dots Y_N, w_L$

➤ Max ratio combination → need large size of buffer  
➤ → need channel information


Construct Data  $K$  bits by encoding  
 $w_1 \dots w_L$  are weights for maximum ration combination

The weights in AWGN channel are

$$w_j = 1 / \sigma_j^2 \quad (j = 1, 2, \dots, L)$$

Which are used for soft-decision decoding in retransmission

Data (+CRC)	K=2000 bits
Error Detection	CRC 24bits
Code rate	1/2
Decoding	Hard decision Viterbi decoding Syndrome decoding
Decoding Error Probability ( $p'$ )	$p' = 10^{-4}$
Channel	AWGN




## 2.2 Another Hybrid ARQ method (2)

**Invertible coding : Retransmitted data packet or redundant packet**

effective for fading channels and jamming

Hybrid ARQ using Invertible code which can decode with only redundant bits. It can be easily performed using RS code with a code rate  $\frac{1}{2}$ .



## 2.2 Conclusions of Error Controlling scheme for BAN

- We proposed the error controlling scheme to **choose hybrid ARQ and FEC only corresponding to medical use and non-medical use, respectively** while transmitted signals have the same channel coding for both medical and non-medical uses.
- The proposed scheme could satisfy the demand of both medical and non-medical simultaneously.
- We can choose a **super orthogonal convolutional code** with much lower code rate but much higher error-correcting capability as well as the same **concatenated code between RS code and convolutional code** as IEEE802.15.4a in option.

## 3. MAC Layer design:

### 3.1 Protocol considering SAR or thermal influence to a body by switching cluster

## 3.1 Background of MAC design

### ■ Integration of medical field and wireless communication technology


- Implementation of network on-body and in-body, **in particular implant BAN inside a body**

- Ex) • Capsule Endoscope
- Cardiac pacemaker
  - etc...

- ✓ Smaller devices
- ✓ Longer-lasting batteries



- Wireless communication devices (nodes) will be able to form a sensor network inside a human body.






# 3.1 Motivation

Focus on thermal influence for a body due to implanted devices


- Problem of wireless communication inside a human body
  - Thermal Influence for a body by electromagnetic wave exposure and circuit heat in a sense of SAR factor

Radiation absorption → Increase of the cell's temperature



## ■ Objective

**Propose a MAC protocol minimizing thermal influence to a human body using a certain thermal propagation model of a body,**




# 3.1 Network Structure of Implant BAN

## ■ Cluster-based communication protocol

- This protocol is more energy efficient than a tree-based protocol.
- Particular nodes (cluster leader) perform long range communication with a receiver outside a body.

## ■ Switching of cluster leader

- **In order to disperse the thermal influence to a body, we switch the access controlling task among multiple FFD (full function device) nodes that are cluster leaders so as to suppress increase of temperature of human tissues .**



**FFD:** Full Function Device

**RFD:** Reduced Function Device

# 3.1 Thermal Propagation Modeling

## 1. Electromagnetic Wave Exposure

■ SAR (Specific Absorption Rate) ... the rate at which radiation energy absorbed by tissue per unit weight

$$SAR = \frac{\sigma}{\rho} E^2 [W / kg]$$

$\sigma$  : electrical conductivity of the tissue [S/m]  
 $\rho$  : density of tissue [kg/m<sup>3</sup>]  
 $E$  : RMS induced electric field [V/m]

Indicator of thermal influence by electromagnetic wave exposure

## 2. Circuit Heat

$$\Delta T = \frac{V \times A}{\rho \times C}$$

$V$ : voltage of leader node [V]  
 $A$ : current of leader node [A]  
 $C$ : specific heat of tissue [J/kgK]

Biologic thermal transport equation

$$\rho c \frac{\partial T}{\partial t} = \kappa \nabla^2 T - \rho \rho_b c_b F (T - T_b) + \rho SAR + \frac{VA}{\rho c}$$


Parameters of blood

## 3.1 Specific Absorption

### 1. Standards for Peak SAR

- We consider about FCC standard in America that is most strict standards than other countries.
- FCC standard  
 $SAR \text{ (per 1g)} < 1.6[\text{W/Kg}]$
- Our value in muscle is  $0.845[\text{W/Kg}]$  that satisfy FCC standards.
- These standards are based on exposure from sources outside the body but we consider sources inside the body.
- So the value of SAR might have to be lower than the standards.

# 3.1 Temperature Characteristic




- ✓ Surrounding medium of the leader is muscle.
- ✓ Consider the temperature in the leader position.
  
- ✓ The more it narrows the switch time, the more the temperature lowers.

Temperature characteristic with or without switching cluster leaders controlling access


# 3.1 Aim of Proposed MAC protocol

## ■ Problem of collecting the information



# 3.1 Access Procedure of Proposed Protocol

## ■ Fundamental access procedure



**Backoff time**  
: a term of carrier sense

Backoff time : Backoff × slot time  
Backoff : integral number randomly generated in the interval [0, CW]  
CW :  $(CW_{\min} + 1) \cdot 2^n - 1$

### 3.1 Adaptive Controlling Back-off Time Algorithm



■ The proposed Algorithm

$$CW_{\min} = \alpha \cdot M$$

**Backoff Coefficient**

Backoff Coefficient Table

$\alpha$	$M$ (The number of RFD node)
4.5	$M < 35$
30	$M \geq 35$




✓ Objective of our algorithm

Derive an appropriate range of backoff time corresponding to the number of RFD nodes to transmit data most efficiently under the restriction of max temperature of human tissues.

Optimum back-off coefficient  $\alpha$  to minimize processing time according to the number of RFD nodes has been pre-calculated and saved.



# 3.1 Flowchart of Proposed Protocol



# 3.1 Performance Evaluation

## Simulation parameters

Data-rate	250kbps
payload	500bits
DATA time	2480 $\mu s$
Slot time	144 $\mu s$
DIFS time	192 $\mu s$
SIFS time	400 $\mu s$
ACK time	352 $\mu s$
Switching interval	10cycles
Number of packets	50packets
RFD nodes	5~40

■ Assumption


assume that range attenuation and packet error are ignored.

✓ Number of packets

··· Data size to transmit of each RFD node in a cycle

✂ extract a referential treatise

# 3.1 Performance Evaluation



## Simulation parameters

Data-rate	250kbps
payload	500bits
DATA time	2480 μs
Slot time	144 μs
DIFS time	192 μs
SIFS time	400 μs
ACK time	352 μs
Switching interval	10cycles
Number of packets	50packets
RFD nodes	5~40


✘extract a referential treatise

### Assumption

We assume that range attenuation and packet error are ignored.

- ✓ Number of packets
- ⋯ Data size to transmit of each RFD node in a cycle

# 3.1 Thermal Influence Evaluation



- ✓ The temperature of leader is saturated after a time caused by cooling effect of blood flow.
- ✓ Proposed protocol can control the thermal influence better than the existing protocols.

## Characteristics of Rising Human Tissue Temperature of Different Protocols

## 3.1 Conclusions of MAC Protocol Considering Thermal Influence to a Body

- We have proposed a novel protocol to minimize thermal influence to a body by switching cluster leaders.
- This protocol may be applicable to any MAC protocol of BAN as a unique approach considering medical purpose of BAN in a sense of SAR.


## 3.2 Positioning or Localization of BAN Nodes in the presence of Non Line of Sights Propagation

## 3.2 Motivation and Objective

- Positioning or Localization system
  - Using an inherent property of UWB, we propose a positioning or localization scheme for BAN.
  - In order to estimate the position of BAN nodes, we focus on Time of Arrival (TOA) positioning method using measurement of distance by UWB .
  - Position information is used to divide the cluster for MAC layer design.
  - It is desirable to reduce the processing quantity for low power consumption.

## 3.2 Assumed environment

- By communicating with each node, distance between nodes is estimated.
- Each position of node is estimated by the distance.





## ▶ 3.2 Positioning Algorithm of TOA Method

– NEWTON algorithm

- (high-speed convergence and a small processing quantity)

- **Compute distance by Observed time**

$$\Delta R_i = c \bullet t_i$$

- **Initial value configuration:  $(x_0, y_0)$**

$$R_i^0 = \sqrt{(X_i - x_0)^2 + (Y_i - y_0)^2}$$

- **Residual matrix from estimation value, linearization**

$$R_i = \frac{\partial R_i}{\partial x} \Delta x + \frac{\partial R_i}{\partial y} \Delta y$$

- **Computing of gradient**

$$\frac{\partial R_i}{\partial x} = - \frac{(X_i - x_0)}{\sqrt{(X_i - x_0)^2 + (Y_i - y_0)^2}}$$

$$\frac{\partial R_i}{\partial y} = - \frac{(Y_i - y_0)}{\sqrt{(X_i - x_0)^2 + (Y_i - y_0)^2}}$$

## 3.2 Positioning Algorithm of TOA Method

- adjust matrix:  $\Delta(X, Y)$

$$\Delta(X, Y) = [\Delta X \ \Delta Y]^T$$

- Residual matrix:  $\Delta R$

$$\Delta R = [\Delta R_1 \ \Delta R_2 \ \dots \ \Delta R_n]^T$$

-Computing of  
gradient matrix **G**

$$G = \begin{bmatrix} \frac{\partial R_1}{\partial x} & \frac{\partial R_1}{\partial y} \\ \frac{\partial R_2}{\partial x} & \frac{\partial R_2}{\partial y} \\ \vdots & \vdots \\ \frac{\partial R_n}{\partial x} & \frac{\partial R_n}{\partial y} \end{bmatrix}$$

- Computing of the Least square solution

$$\Delta(X, Y) = (G^T G)^{-1} G^T \Delta R \longrightarrow \text{Addition to } (x_0, y_0) \text{ ,and repeat}$$

## 3.2 Mitigation of Influence of Non Line of Sight (NLOS) Paths

- A barrier or undesired object between transmission and reception nodes makes arrival time longer due to indirect path with non-arrival of direction wave.
- This effect of non line of sight (NLOS) leads to degradation of positioning accuracy.
- Using temporarily estimated position of all nodes, NLOS paths can be selected and removed or mitigated in a list of available paths for more accurate positioning.


## 3.2 Mitigation of NLOS Paths influence

Estimated position by all nodes:  $[X, Y]_{temp} = [X_{temp}, Y_{temp}]$

NLOS influenced value:  $\Delta L_i = \Delta R_i - \sqrt{(X_i - X_{temp})^2 + (Y_i - Y_{temp})^2}$


- Remove largest NLOS path node as worst influenced node of NLOS paths.
- Excluding removed NLOS path node, positioning process is performed by information of other LOS nodes.
- These processes are repeated at  $Nd$  times, where  $Nd$  is the number of removed nodes.

# 3.2 Simulation Results



Field	10x10 [m]
Node number	9
Deleted number $N_d$	1,2,3,4
Trial number	50000
Distance measurement error	0.2[m] (Gaussian)
Node position error	0.2[m] (Gaussian)
Average NLOS Delay	2 [m](Exponential)

# 3.2 Simulation Results




Field	10x10 [m]
Node number	9
Deleted number $N_d$	1,2,3,4
Trial number	50000
Distance measurement error	0.2[m] (Gaussian)
Node position error	0.2[m] (Gaussian)
NLOS rate	0.3

Optimal  $N_d$  is determined by each NLOS rate, distance measurement error, node position error or node number.

## 3.2 Decrease of Processing Complexity

- Number of computation on node information (approximation number)
  - All search algorithm ... 2500 times
  - Partial Filter algorithm ... 400 times
  - NEWTON algorithm ... 10 times
- NEWTON algorithm is effective to reduce processing complexity and power consumption.



## 3.2 Concluding Remarks of BAN Nodes Positioning or Localization

- We proposed an optional function of BAN standard as the positioning system robust against performance degradation due to non line of sight (NLOS) propagation paths.
- We focused on NEWTON algorithm (TOA) in perspective of low consumption but we have published many other alternative choices applicable for BAN node localization.
- NLOS influence in accuracy of positioning can be mitigated by removing NLOS path nodes.



## 4. Concluding Remarks

- We have proposed some useful technologies in PHY and MAC of BAN satisfying requirement of both medical and non-medical uses considering trade-off between reliability and efficiency.
- For PHY, a pulsed chirp UWB using hopping and combined hybrid ARQ and FEC for medical and non-medical uses have been proposed.
- For MAC, a control scheme of thermal influence by switching cluster leaders has been proposed.
- In addition of our proposal in March, a positioning or localization scheme of BAN nodes has been introduced using UWB inherent property.
- This presentation could be useful to make a complete standard of BAN more attractive for both medical and non-medical uses by introducing these key schemes applicable to BAN.

# Reference

1. Hideki Mochizuki, Makoto, Kawasaki, Shun Nagamine, Igor Dotlic, Ryuji Kohno, "Performance Analysis of Pulsed Chirp UWB Schemes used Hopping Sequence for Wearable Wireless Body Area Network," 2009 International Symposium on Medical Information Communications Technology (ISMICT2009), Montreal (Canada), Feb.24-26, 2009
2. Shun Nagamine, Ryuji Kohno, "Design of Communication Model Suitable for Implanted Body Area Networks," 2009 International Symposium on Medical Information Communications Technology (ISMICT2009), Montreal (Canada), Feb.24-26, 2009
3. Haruka Suzuki, Ryuji Kohno, "Error-Controlling Scheme for UWB Wireless Communications for both Medical and Non-Medical Uses," IEICE Technical Conference on Medical ICT, Tokyo (Japan), April 23, 2009
4. Makoto Kawasaki, Ryuji Kohno, "Position Estimation Method of Medical Implanted Devices Using Estimation of Propagation Velocity inside Human Body, "IEICE TRANS. Special Section on Medical Information and Communications Technologies, Vol.E92-B, No.02 Feb. 2009
5. Matti Hamalainen, Attaphongse Taparugssanagorn, Jari Linatti, Ryuji Kohno, "Exploitation of Wireless Technology in Remote Care Processes," IEICE Transactions on Communications, Vol. E92-B, No.2, pp.373 – 378, Feb. 2009.
6. Koji Enda, Ryuji Kohno, "Iterative Delay Compensation Algorithm to Mitigate NLOS Influence for Positioning," International Symposium on Information Theory and its Applications 2008 (ISITA2008) Auckland, New Zealand, pp.370-375, Dec. 2008
7. Koji Enda, Ryuji Kohno, "Routing Algorithm to Decrease Transmission Delay and Traffic Quantity using Priority Area in Sensor Network of Hospital," The Third International Symposium on Medical Information & Communication Technology (ISMICT2009), Montreal, Canada, Feb, 2009